

Hasil Uji Data Statistik SPSS Versi 25

a. Hasil Uji Normalitas berdasarkan Metode Otomatis dan Manual

	METODE	Kolmogorov-Smirnov ^a			Shapiro-Wilk		
		Statistic	df	Sig.	Statistic	df	Sig.
LIM	OTOMATIS	.088	36	.200*	.975	36	.564
	MANUAL	.082	36	.200*	.967	36	.351
MID	OTOMATIS	.244	36	.000	.586	36	.000
	MANUAL	.190	36	.002	.817	36	.000
GRAN	OTOMATIS	.137	36	.085	.967	36	.345
	MANUAL	.130	36	.130	.970	36	.420

b. Hasil Uji Normalitas berdasarkan Lama Penyimpanan

	PENUNDAAN	Kolmogorov-Smirnov ^a			Shapiro-Wilk		
		Statistic	df	Sig.	Statistic	df	Sig.
LIM	0 JAM	.100	24	.200*	.958	24	.399
	24 JAM	.085	24	.200*	.982	24	.936
	60 JAM	.142	24	.200*	.940	24	.161
MID	0 JAM	.129	24	.200*	.981	24	.915
	24 JAM	.138	24	.200*	.973	24	.738
	60 JAM	.249	24	.000	.763	24	.000
GRAN	0 JAM	.156	24	.134	.954	24	.327
	24 JAM	.079	24	.200*	.984	24	.958
	60 JAM	.174	24	.060	.928	24	.088

c. Hasil Uji Normalitas berdasarkan Jenis Antikoagulan

	JENIS ANTIKOAGULAN	Kolmogorov-Smirnov ^a			Shapiro-Wilk		
		Statistic	df	Sig.	Statistic	df	Sig.
LIM	Na2EDTA	.108	36	.200*	.962	36	.251
	K3EDTA	.089	36	.200*	.965	36	.303
MID	Na2EDTA	.263	36	.000	.678	36	.000
	K3EDTA	.109	36	.200*	.944	36	.068
GRAN	Na2EDTA	.150	36	.041	.935	36	.035
	K3EDTA	.104	36	.200*	.980	36	.731

d. Uji Non Parametrik *Kruskal-Wallis* dan *Mann-Whitney U*

Kruskal-Wallis Test

Ranks			
	PENUNDAAN	N	Mean Rank
MID	0 JAM	24	38.67
	24 JAM	24	34.06
	60 JAM	24	36.77
	Total	72	

Test Statistics^{a,b}

MID	
Kruskal-Wallis H	.590
df	2
Asymp. Sig.	.745

a. Kruskal Wallis Test

b. Grouping Variable:
PENUNDAAN

**NON-PARAMETRIK LAMA PENYIMPANAN
PADA PARAMETER MID CELL**

Mann-Whitney Test

Ranks			
	METODE	N	Mean Rank
MID	OTOMATIS	36	41.31
	MANUAL	36	31.69
	Total	72	

Test Statistics^a

MID	
Mann-Whitney U	475.000
Wilcoxon W	1141.000
Z	-1.954
Asymp. Sig. (2-tailed)	.051

a. Grouping Variable:
METODE

**NON-PARAMETRIK METODE PADA
PARAMETER MID CELL**

Mann-Whitney Test

Ranks				
	JENIS ANTIKOAGULAN	N	Mean Rank	Sum of Ranks
MID	Na2EDTA	36	39.53	1423.00
	K3EDTA	36	33.47	1205.00
	Total	72		
GRAN	Na2EDTA	36	35.88	1291.50
	K3EDTA	36	37.13	1336.50
	Total	72		

Test Statistics^a

	MID	GRAN
Mann-Whitney U	539.000	625.500
Wilcoxon W	1205.000	1291.500
Z	-1.231	-.254
Asymp. Sig. (2-tailed)	.218	.800

a. Grouping Variable: JENIS
ANTIKOAGULAN

**NON-PARAMETRIK JENIS ANTIKOAGULAN
PADA PARAMETER MID CELL DAN
GRANULOSIT**

e. Hasil Uji Statistik *One-Way ANOVA*

1. Parameter Limfosit berdasarkan Lama Penyimpanan

LIM

	Sum of Squares	df	Mean Square	F	Sig.
Between Groups	43.837	2	21.918	.387	.681
Within Groups	3909.922	69	56.666		
Total	3953.759	71			

2. Parameter Limfosit berdasarkan Lama Jenis Antikoagulan

LIM

	Sum of Squares	df	Mean Square	F	Sig.
Between Groups	1.280	1	1.280	.023	.881
Within Groups	3952.479	70	56.464		
Total	3953.759	71			

3. Parameter Linfosit berdasarkan Jenis Metode

LIM

	Sum of Squares	df	Mean Square	F	Sig.
Between Groups	11.361	1	11.361	.202	.655
Within Groups	3942.399	70	56.320		
Total	3953.759	71			

4. Parameter Granulosit berdasarkan Lama Penyimpanan

GRAN

	Sum of Squares	df	Mean Square	F	Sig.
Between Groups	99.924	2	49.962	.651	.525
Within Groups	5298.428	69	76.789		
Total	5398.353	71			

5. Parameter Granulosit berdasarkan Metode

GRAN

	Sum of Squares	df	Mean Square	F	Sig.
Between Groups	.934	1	.934	.012	.913
Within Groups	5397.419	70	77.106		
Total	5398.353	71			

f. Hasil Uji Statistik Nilai Rata-Rata dan Standar Deviasi terhadap *Differential Leukocyte Count (DLC)*

PENUNDAAN	METODE	JENIS ANTIKOAGULAN		LIM	MID	GRAN	
0 JAM	OTOMATIS	Na2EDTA	Mean	35.300	7.750	56.950	
			N	6	6	6	
			Std. Deviation	7.6663	.7556	8.2803	
		K3EDTA	Mean	35.417	7.200	57.383	
			N	6	6	6	
			Std. Deviation	7.5959	.8922	8.4072	
		Total	Mean	35.358	7.475	57.167	
			N	12	12	12	
			Std. Deviation	7.2763	.8390	7.9589	
		MANUAL	Na2EDTA	Mean	37.500	6.333	56.167
				N	6	6	6
				Std. Deviation	8.7350	.5164	9.0646
	K3EDTA		Mean	36.500	6.667	56.667	
			N	6	6	6	
			Std. Deviation	8.1179	2.4221	8.2138	
	Total		Mean	37.000	6.500	56.417	
			N	12	12	12	
			Std. Deviation	8.0566	1.6787	8.2513	
	Total		Na2EDTA	Mean	36.400	7.042	56.558
				N	12	12	12
				Std. Deviation	7.9194	.9634	8.2874
			K3EDTA	Mean	35.958	6.933	57.025
				N	12	12	12
				Std. Deviation	7.5167	1.7624	7.9331
		Total	Mean	36.179	6.988	56.792	
			N	24	24	24	
			Std. Deviation	7.5543	1.3901	7.9375	
24 JAM	OTOMATIS	Na2EDTA	Mean	36.567	6.967	56.467	
			N	6	6	6	
			Std. Deviation	4.4230	.6154	4.9168	
		K3EDTA	Mean	33.917	6.583	59.500	
			N	6	6	6	
			Std. Deviation	9.1973	.4491	9.4839	
		Total	Mean	35.242	6.775	57.983	
			N	12	12	12	

60 JAM	MANUAL	Na2EDTA	Std. Deviation	7.0183	.5512	7.3744
			Mean	36.167	6.667	57.167
			N	6	6	6
		K3EDTA	Std. Deviation	5.2313	1.9664	6.5549
			Mean	35.333	6.500	58.167
			N	6	6	6
		Total	Std. Deviation	8.4538	1.0488	8.6120
			Mean	35.750	6.583	57.667
			N	12	12	12
	Total	Na2EDTA	Std. Deviation	6.7167	1.5050	7.3154
			Mean	36.367	6.817	56.817
			N	12	12	12
		K3EDTA	Std. Deviation	4.6233	1.3979	5.5365
			Mean	34.625	6.542	58.833
			N	12	12	12
		Total	Std. Deviation	8.4547	.7704	8.6649
			Mean	35.496	6.679	57.825
			N	24	24	24
		OTOMATIS	Std. Deviation	6.7232	1.1128	7.1853
			Mean	36.217	9.050	54.733
			N	6	6	6
60 JAM	OTOMATIS	Na2EDTA	Std. Deviation	8.1160	4.4801	12.2127
			Mean	38.317	6.850	54.833
			N	6	6	6
		K3EDTA	Std. Deviation	9.9381	1.2566	11.1315
			Mean	37.267	7.950	54.783
			N	12	12	12
		Total	Std. Deviation	8.7199	3.3408	11.1410
			Mean	37.167	8.000	54.833
			N	6	6	6
	MANUAL	Na2EDTA	Std. Deviation	7.4140	4.0497	10.6286
			Mean	37.833	6.667	55.500
			N	6	6	6
		K3EDTA	Std. Deviation	9.4322	2.9439	12.0789
			Mean	37.500	7.333	55.167
			N	12	12	12
		Total	Std. Deviation	8.0960	3.4466	10.8530
			Mean	36.692	8.525	54.783
			N	12	12	12
	Total	Na2EDTA	Std. Deviation	7.4278	4.1083	10.9154
			Mean	36.692	8.525	54.783
			N	12	12	12

Total	OTOMATIS	K3EDTA	Mean	38.075	6.758	55.167
			N	12	12	12
			Std. Deviation	9.2410	2.1602	11.0798
		Total	Mean	37.383	7.642	54.975
			N	24	24	24
			Std. Deviation	8.2297	3.3344	10.7580
		Na2EDTA	Mean	36.028	7.922	56.050
			N	18	18	18
			Std. Deviation	6.5357	2.6390	8.4913
Total	OTOMATIS	K3EDTA	Mean	35.883	6.878	57.239
			N	18	18	18
			Std. Deviation	8.6273	.9085	9.3564
		Total	Mean	35.956	7.400	56.644
			N	36	36	36
			Std. Deviation	7.5435	2.0159	8.8264
	MANUAL	Na2EDTA	Mean	36.944	7.000	56.056
			N	18	18	18
			Std. Deviation	6.8554	2.5668	8.4260
		K3EDTA	Mean	36.556	6.611	56.778
			N	18	18	18
			Std. Deviation	8.2264	2.1458	9.2645
		Total	Mean	36.750	6.806	56.417
			N	36	36	36
			Std. Deviation	7.4656	2.3400	8.7354
	Total	Na2EDTA	Mean	36.486	7.461	56.053
			N	36	36	36
			Std. Deviation	6.6174	2.6079	8.3370
		K3EDTA	Mean	36.219	6.744	57.008
			N	36	36	36
			Std. Deviation	8.3149	1.6296	9.1795
		Total	Mean	36.353	7.103	56.531
			N	72	72	72
			Std. Deviation	7.4624	2.1891	8.7197

g.